



Radiation Measurement Instruments
Засоби радіаційного контролю

DECLARATION OF CONFORMITY ДЕКЛАРАЦІЯ ВІДПОВІДНОСТІ



We, manufacturer: **PE "SPPE "Sparing-Vist Center"**

33 Volodymyr Velyky Str., Lviv, 79026 Ukraine

declare under our sole responsibility, that the product
Search Alarm Radiometer

DKS-02PN "CADMIUM". Trade mark: **ECOTEST**
to which this declaration relates, is in conformity
with the following standards:

Ми, виробник: **ПП "НВП "Спаринг-Віст Центр"**

заявляємо під свою відповідальність, що виріб

Дозиметр-сигналізатор пошуківий
ДКС-02ПН "КАДМІЙ". Торгова марка: **ЕКОТЕСТ**
згідно з даною декларацією відповідає
наступним стандартам:

IEC 61326-1	Electrical equipment, control and laboratory use - EMC requirements.
CISPR 11	Conducted Emissions - (Group 1, Class B); Radiated Emissions - (Group 1, Class B).
IEC 61000-3-2	Limits for Harmonic Current Emissions - (Class A).
IEC 61000-3-3	Limitation of Voltage Fluctuations and Flicker.
IEC 61000-4-2	Immunity to Electrostatic Discharges: ESD Air 8 kV - (Criterion A); ESD Contact 4 kV - (Criterion B).
IEC 61000-4-3	Immunity to Radio Frequency Electromagnetic Fields - (Criterion A): 10 V/m 80 - 1000 MHz; 3 V/m 1400 - 2000 MHz; 1 V/m 2000 - 2700 MHz.
IEC 61000-4-4	Immunity to Electrical Fast Transients - (Criterion B), Power Ports: ± 2 kV / 5 kHz.
IEC 61000-4-5	Immunity to Surge - (Criterion A), Power Ports: ± 1 kV / Line to Line.
IEC 61000-4-6	Immunity to Conducted Disturbances - (Criterion A), Power Ports: 3V
IEC 61000-4-8	Immunity to Power-Frequency Magnetic Fields - (Criterion A), 30 A/m, 50 and 60 Hz.
IEC 61000-4-11	Immunity to Voltage Dips, Interruptions and Variations: Voltage Dips = 30%, Period = 25, Cycles = 30 (Criterion A); Voltage Dips = 60%, Period = 10, Cycles = 12 (Criterion A); Voltage Dips = 100%, Period = 1 (Criterion A); Voltage Interruption = 100%, Period = 250, Cycles = 300 (Criterion B).
IEC 61010 -1	Safety requirements for electrical equipment for measurement, control and laboratory use.

in accordance with

EMC-directive 2014/30/EU and RoHS-directive 2011/65/EU

The tests for compliance with stated
above standards have been conducted at:

у відповідності з умовами

випробовування на відповідність

вищезазначеним стандартам проводились:

UkrTEST of Ukmetrteststandart
4 Metrologichna str., Kyiv, 03143, Ukraine

Report Reference No 0024-5-2018

Lviv, July, 2018
Y. Storonsky
Director

Manufacturer / Виробник:

м. Львів, липень 2018р.
Ю. Сторонський
Директор

