



Radiation Measurement Instruments
Засоби радіаційного контролю

DECLARATION OF CONFORMITY ДЕКЛАРАЦІЯ ВІДПОВІДНОСТІ



We, manufacturer: PE "SPPE "Sparing-Vist Center"

33 Volodymyr Velyky Str., Lviv, 79026 Ukraine

declare under our sole responsibility, that the product
EPD-27 «DoseG» («DoseGX») Electronic Personal
Dosimeter with **USB/IrDA** Adapter.

Trade mark: **ECOTEST** to which this declaration
relates, is in conformity with the following standards:

Ми, виробник: ПП "НВП "Спаринг-Віст Центр"

заявляємо під свою відповідальність, що виріб
Дозиметр електронний персональний **ЕПД-27**
«DoseG» («DoseGX») з адаптером **USB/IrDA**.
Торгова марка: **ЕКОТЕСТ** згідно з даною
декларацією відповідає наступним стандартам:

IEC 61326-1:2012 (EN 61326-1:2013)
CISPR 11:2009 +A1:2000

Electrical equipment, control and laboratory use – EMC requirements.
Conducted Emissions - (Group 1, Class B);
Radiated Emissions - (Group 1, Class B).

IEC 61000-4-2:2008

Immunity to Electrostatic Discharges: ESD Air 8 kV - (Criterion A);
ESD Contact 4 kV - (Criterion B).

IEC 61000-4-3:2006 +A1:2007 +A2:2010

Immunity to Radio Frequency Electromagnetic Fields - (Criterion A):
10 V/m 80 – 1000 MHz;
3 V/m 1400 – 2000 MHz;
1 V/m 2000 – 2700 MHz.

IEC 61000-4-8:2009

Immunity to Power-Frequency Magnetic Fields - (Criterion A), 30 A/m,
50 and 60 Hz.

IEC 61010-1:2010 +A1:2016 (EN 61010-1:2010 +A1:2019)

Safety requirements for electrical equipment for
measurement, control and laboratory use.

in accordance with

EMC-directive 2014/30/EU and RoHS-directive 2011/65/EU

The tests for compliance with stated
above standards have been conducted at:

у відповідності з умовами

Випробовування на відповідність

вищезазначеним стандартам проводились:

UkrTEST of Ukrmetrteststandart

4 Metrologichna str., Kyiv, 03143, Ukraine

Report Reference No 0507-5-2021 / 11.06.2021, 0510-1-2021 / 07.05.2021

Manufacturer/Виробник:

Lviv, September, 2021

Y. Storonskyi
Director

м. Львів, вересень 2021р.

Ю. Сторонський
Директор

