



Radiation Measurement Instruments
Засоби радіаційного контролю

DECLARATION OF CONFORMITY

ДЕКЛАРАЦІЯ ВІДПОВІДНОСТІ



We, manufacturer: **PE "SPPE "Sparing-Vist Center"**

33 Volodymyr Velyky Str., Lviv, 79026 Ukraine

declare under our sole responsibility, that the product
Search Dosimeter-Radiometer

MKS-11 "SPECTRA". Trade mark: **ECOTEST**
to which this declaration relates, is in conformity
with the following standards:

Ми, виробник: **ПП "НВП "Спаринг-Віст Центр"**

заявляємо під свою відповідальність, що виріб
Дозиметр-радіометр пошуковий

МКС-11 "СПЕКТРА". Торгова марка: **ЕКОТЕСТ**
згідно з даною декларацією відповідає
наступним стандартам:

- | | |
|-----------------------|---|
| IEC 61326-1 | Electrical equipment, control and laboratory use – EMC requirements. |
| CISPR 11 | Conducted Emissions - (Group 1, Class B); Radiated Emissions - (Group 1, Class B). |
| IEC 61000-3-2 | Limits for Harmonic Current Emissions - (Class A). |
| IEC 61000-3-3 | Limitation of Voltage Fluctuations and Flicker. |
| IEC 61000-4-2 | Immunity to Electrostatic Discharges: ESD Air 8 kV - (Criterion A); ESD Contact 4 kV - (Criterion B). |
| IEC 61000-4-3 | Immunity to Radio Frequency Electromagnetic Fields - (Criterion A):
10 V/m 80 – 1000 MHz;
3 V/m 1400 – 2000 MHz;
1 V/m 2000 – 2700 MHz. |
| IEC 61000-4-4 | Immunity to Electrical Fast Transients - (Criterion B), Power Ports: ± 2 kV / 5 kHz. |
| IEC 61000-4-5 | Immunity to Surge - (Criterion A), Power Ports: ± 1 kV / Line to Line. |
| IEC 61000-4-6 | Immunity to Conducted Disturbances - (Criterion A), Power Ports: 3V |
| IEC 61000-4-8 | Immunity to Power-Frequency Magnetic Fields - (Criterion A), 30 A/m, 50 and 60 Hz. |
| IEC 61000-4-11 | Immunity to Voltage Dips, Interruptions and Variations:
Voltage Dips = 30%, Period = 25, Cycles = 30 (Criterion A);
Voltage Dips = 60%, Period = 10, Cycles = 12 (Criterion A);
Voltage Dips = 100%, Period = 1 (Criterion A);
Voltage Interruption = 100%, Period = 250, Cycles = 300 (Criterion B). |
| IEC 61010 -1 | Safety requirements for electrical equipment for measurement, control and laboratory use. |

in accordance with

у відповідності з умовами

EMC-directive 2014/30/EU and RoHS-directive 2011/65/EU

The tests for compliance with stated

Випробовування на відповідність

above standards have been conducted at:

вищезазначеним стандартам проводились:

UkrTEST of Ukmetteststandart

4 Metrologichna str., Kyiv, 03143, Ukraine

Report Reference No 0022-5-2018

Manufacturer / Виробник:

Lviv, July, 2018

Y. Storonsky

Director

м. Львів, липень 2018р.

Ю. Сторонський

Директор

